




Date: 27 May 2014  
Calibration of: M4260  
Serial Number: 3715

- Measurement Data on Receipt: **in tolerance**

- Detailed Calibration Test Results:

|                                 |            |                   |                                      |
|---------------------------------|------------|-------------------|--------------------------------------|
| System calibration              | before     | actual            | calibration uncertainty <sup>1</sup> |
| Sensitivity @ 1 kHz, 114 dB SPL | 26.1 mV/Pa | <b>26.3 mV/Pa</b> | ±2.85%                               |
| Frequency response              | Class 2    | acc. IEC 61672    |                                      |



- Test Conditions: Temperature: **24.1°C** ±0.5 °C  
Relative Humidity: **40.3%** ±2%  
Air Pressure: **96.01 kPa** ±0.25 kPa

- Calibration Equipment Used:

- Norsonic Sound Calibrator, Type 1251, S/No. 30930  
Last Calibration: 3. Dec. 2012, Next Calibration: 3. Dec. 2014  
Calibrated by Metas, Switzerland
- NTi Audio FX100, S/No. 10402  
Last Calibration: 17. Aug. 2013, Next Calibration: 17. Aug. 2014  
Calibrated by NTi Audio meeting product specifications
- MTG MV203, S/No. 0630 / Mic Capsule, MK221 S./No. 13164  
Last Calibration: 26. Jun. 2012, Next Calibration: 30. Jun. 2014  
Calibrated by MTG, Germany

<sup>1</sup> The expanded uncertainty assigned to the measurement results is obtained by multiplying the standard uncertainty by the coverage factor  $k = 2$ . The value of the measurand lies within the assigned range of values with a probability of 95%.

Calibration of: XL2  
 Serial Number: A2A-06530-E0  
 Date: 27 May 2014

- Measurement Data on Receipt: **in tolerance**

- Detailed Calibration Test Results:

|                                  |        | reference | before  | actual           | unit | actual error | XL2 tolerance | calibration uncertainty <sup>2</sup> |
|----------------------------------|--------|-----------|---------|------------------|------|--------------|---------------|--------------------------------------|
| RMS Level @ 1kHz, XLR Input      |        | 0.1       | 0.100   | <b>0.100</b>     | V    | ≤0.1%        | ±0.5%         | ±0.10%                               |
|                                  |        | 1         | 1.000   | <b>1.000</b>     | V    | ≤0.1%        | ±0.5%         | ±0.09%                               |
|                                  |        | 10        | 9.991   | <b>9.990</b>     | V    | -0.1%        | ±0.5%         | ±0.09%                               |
| Flatness, XLR Input <sup>1</sup> | 20 Hz  | 1         | 0.997   | <b>0.997</b>     | V    | -0.3%        | ±1.1%         | ±0.09%                               |
|                                  | 20 kHz | 1         | 1.005   | <b>1.004</b>     | V    | 0.4%         | ±1.1%         | ±0.09%                               |
| Frequency                        |        | 1000      | 1000.00 | <b>1000.01</b>   | Hz   | ≤0.003%      | ±0.003%       | ±0.01%                               |
| Residual Noise                   | XLR    |           | < 2 uV  | <b>&lt; 2 uV</b> |      |              | <2 uV         | ±0.50%                               |
| THD+N @ 0 dBu, 1 kHz, XLR Input  |        |           | -99.2   | <b>-99.2</b>     | dB   |              | typ. -100 dB  | ±0.50%                               |

- Test Conditions: Temperature: **24** °C  
 Relative Humidity: **37** %

- Calibration Equipment Used:

- Agilent Multimeter, Typ 34401A, Serial No. US 3611 3653  
 Last calibration: 12.08.2013. Next calibration due: 12.08.2014  
 Calibrated by ELCAL to the national standards maintained at Swiss Federal Office of Metrology. SCS 002
- FX100 Audio Analyzer, Serial No. 10408  
 Last Calibration: 08.05.2013, Next Calibration: 08.05.2014  
 Manufacturer calibration based on Agilent 34410, Serial No. MY47014254,  
 Last Calibration: 05.06.2013, Next Calibration: 05.06.2014  
 which is calibrated by ELCAL to national standards maintained at Swiss Federal Office of Metrology. SCS 002

<sup>1</sup> The specified tolerance  $\pm 0.1$  dB @ 1V =  $\pm 1.1$  %

<sup>2</sup> The expanded uncertainty assigned to the measurement results is obtained by multiplying the standard uncertainty by the coverage factor  $k = 2$ . The value of the measurand lies within the assigned range of values with a probability of 95%.

